Autoclave Maintenance and Run Log Year _____

	JAN	FEB	MAR	APR	MAY	JUN	JUL	AUG	SEP	OCT	NOV	DEC
Monthly Spore Testing Results												
Monthly Cleaning												
Annual Calibration												

Sterility is EVENT related, not time related; Pack is considers sterile unless an event causes contamination (example: punctured, torn, cracked packs= unsterile; evidence of water damage or yellowed packs= unsterile) Have Process to routine evaluation of sterile packs.

Log Process: Write date and Load # on Pack (if more than 1 load is run in the same day, write date and load #1 and then date and load #2 etc)

Date	Time	Load#	Item(s)	Temperature(250-254 Degrees)	Steam Pressure (15-17 psi)	Duration of Run (30 Minutes)	Person Responsible

Sterility is EVENT related, not time related; Pack is considers sterile unless an event causes contamination (example: punctured, torn, cracked packs= unsterile; evidence of water damage or yellowed packs= unsterile) Have Process to routine evaluation of sterile packs.

Log Process: Write date and Load # on Pack (if more than 1 load is run in the same day, write date and load #1 and then date and load #2 etc)

Date	Time	Load#	Item(s)	Temperature(250-254 Degrees)	Steam Pressure (15-17 psi)	Duration of Run (30 Minutes)	Person Responsible